

205.10 Radon Emanation (encapsulated solution form)

These SRMs are intended for the calibration of radon-222 measuring instruments.

[Click here for more information.](#)

[Technical Contact: ronald.colle@nist.gov](mailto:ronald.colle@nist.gov)

PLEASE NOTE:
For specific values and uncertainties, the certificate is the only official source.

| SRM | Radionuclide | Radium 226 Activity (Bq) | Time of Calibration (month/year) |
|------|------------------------------|--------------------------|-------------------------------------|
| 4971 | Radon-222 Emanation Standard | 5.192 ± 0.062 | 9/03 |
| 4972 | Radon-222 Emanation Standard | 49.55 ± 0.59 | 9/03 |
| 4973 | Radon-222 Emanation Standard | 517.7 ± 6.2 | 9/03 |
| 4974 | Radon-222 Emanation Standard | 5761 ± 75 | 5/05 |